

**Resistin (Mouse)
ELISA Kit Protocol**

(Cat. No.: EK-28-35)

PHOENIX PHARMACEUTICALS, INC.
330 Beach Road, Burlingame, California 94010
Tel: 650-558-8988 800-988-1205 Fax: 650-558-1686
E-Mail: Info@PhoenixPeptide.com
www.PhoenixPeptide.com

Resistin (Mouse) ELISA KIT PROTOCOL

TABLE OF CONTENTS

Introduction and Protocol Overview	1
List of Components	3
Material Required but Not Supplied	3
Reagent Preparation	4
Mouse Resistin Standard Preparation	4
Mouse Resistin ELISA Protocol	5
Additional Recommended Procedural Note	7
Summary of Assay Protocol	8
Calculation of Results	9
Storage	10
Note	10

Resistin (Mouse) ELISA Kit Protocol

INTRODUCTION AND PROTOCOL OVERVIEW

Resistin, a novel signalling molecule isolated in mice has been suggested to be the putative hormone thought to link obesity with type 2 diabetes. Resistin is a member of a newly discovered, small family of cysteine-rich secretory proteins (1,2). It is expressed only in adipose tissue and is markedly induced during the differentiation of 3T3-L1 cells to mature adipocytes (1). Initial studies in rodents suggested that resistin is upregulated in both genetic and diet-induced obesity and is downregulated by thiazolidinediones (TZDs). Resistin administration also impaired glucose tolerance, whereas immunoneutralization of this protein improved insulin action. Hence, this molecule has been proposed as a link between obesity and insulin resistance.

Phoenix Pharmaceutical's Mouse Resistin ELISA Kit is designed to measure the concentration of Mouse Resistin from Mouse serum/plasma, or conditioned medium.

The immunoplate in this kit is pre-coated with Anti-Mouse Resistin Capture Antibody and the non-specific binding sites are blocked. The Mouse Resistin in the sample or in the standard solution can bind to the capture antibody immobilized in the wells. After washing procedure, the biotinylated anti-Mouse Resistin Detection Antibody which can bind to the Mouse Resistin trapped in the wells is added. The enzyme-substrate reaction is terminated by the addition of a stop solution. The intensity of the color is directly proportional to the amount of Mouse Resistin in the standard solutions or samples. A standard curve of Mouse Resistin with known concentration can be established accordingly. The Mouse Resistin with unknown concentration in samples can be determined by extrapolation to this standard curve.


ASSAY CONDITIONS:

Plasma, serum, culture media, tissue homogenate, CSF, urine or any biological fluid can be assayed as long as the level of the sample is high enough for the sensitivity of the kit to detect it.

Resistin (Mouse) ELISA Kit Protocol

CAUTION: Phoenix Pharmaceuticals guarantees that its products conform to the information contained in this publication. The purchaser must determine the suitability of the product for their particular needs and establish optimum sample concentrations.

Assay Principle



Resistin (Mouse) ELISA Kit Protocol

LIST OF COMPONENTS

Store all components at 4°C. DO NOT FREEZE.

1. 20x Assay Buffer concentrate (50ml).....Catalog no. EK-BUF
2. 96 Well Anti-Mouse Resistin.....Catalog no. EK-Plate-028-35
Capture Antibody-Coated Plate (1Plate)
3. Mouse Resistin Standard.....Catalog no. EK-S-028-35
(100ng/ml)
4. Biotinylated Anti-Mouse Resistin.....Catalog no. EK-B-028-35
Detection Antibody (1 vial)
5. Mouse Resistin.....Catalog no. EK-PC-028-35
Positive Control (2 vials)
6. Streptavidin-horseradish peroxidase.....Catalog no. EK-SA-HRP
(SA-HRP) (30µl)
7. Substrate Solution (TMB) (12ml).....Catalog no. EK-SS
8. Stop Solution 2N HCl (15ml).....Catalog no. EK-HCl
9. Acetate plate sealer (APS) (3 pieces).....Catalog no. EK-APS
10. Assay Diagram (1 sheet)

MATERIAL REQUIRED BUT NOT SUPPLIED

- Micropipettor (s) and disposable pipette tips
- Multi-channel pipette capable of dispensing 50-100µl
- Solution reservoir (recommended)
- Microtiter plate washer (recommended)
- Orbital plate shaker capable of 300-500rpm (recommended)
- Microtiter plate reader capable of absorbance measurement between 450nm-650nm
- Well-closed containers (15ml tubes or more in capacity)
- Absorbent material for blotting

Resistin (Mouse) ELISA Kit Protocol

REAGENT PREPARATION

Note: The kit should be equilibrated to room temperature (20-23°C) before opening any vials and starting the assay. It is highly recommended that the solutions be used as soon as possible after rehydration.


1. **1x Assay buffer:** Dilute the **20x** assay buffer concentrate with 950ml of distilled water. This assay buffer will be used to wash the plate and reconstitute all of the other compounds in this kit. If crystals are observed in the 20x Assay buffer, warm the bottle in a 37°C water bath for approximately 30 minutes or until the crystals disappear. After preparation, store **1x** Assay buffer at 4°C.
2. **Biotinylated anti-mouse Resistin Detection Antibody:** Rehydrate biotinylated anti-Mouse Resistin detection antibody with 100µl of **1x** assay buffer, vortex (centrifuge the tube to dislodge powder from the cap or walls). Dilute biotinylated anti-Mouse Resistin Detection Antibody to 1:100 and mix thoroughly before use.
3. **Streptavidin-Horseradish Peroxidase (SA-HRP):** Centrifuge the HRP vial (30µl) provided in this kit (3,000-5,000 rpm, 5 seconds) and dilute SA-HRP with **1x** assay buffer to 1:2000 before use. Vortex thoroughly.
4. **Mouse Resistin Positive Control:** Rehydrate Mouse Resistin Positive Control with 250µl of **1x** assay buffer (centrifuge the tube to dislodge powder from cap or walls). Vortex thoroughly.

MOUSE RESISTIN STANDARD PREPARATION

1. Rehydrate recombinant Resistin standard with 1ml **1x** assay buffer, vortex. Allow the solution to sit at least 10 minutes at room temperature (20-23°C) to completely dissolve in solution. Vortex and centrifuge before use. The concentration of this stock solution is 100ng/ml.
2. Prepare Mouse Resistin standard solutions as follows:

Resistin (Mouse) ELISA Kit Protocol

Standard No.	Std. volume	Assay Buffer	Concentrations
Stock	Powder	1000 μ l	100ng/ml
#1	10 μ l of stock	990 μ l	1ng/ml
#2	500 μ l of #1	500 μ l	0.5ng/ml
#3	500 μ l of #2	500 μ l	0.25ng/ml
#4	500 μ l of #3	500 μ l	0.125ng/ml
#5	500 μ l of #4	500 μ l	0.062ng/ml
#6	500 μ l of #5	500 μ l	0.031ng/ml
#7	500 μ l of #6	500 μ l	0.015ng/ml



HUMAN IL-6 ELISA PROTOCOL

1. Thoroughly read this protocol before performing an assay.
2. Remove Capture Antibody-Coated Plate from its zip-lock foil pouch. Remove any unneeded strips from the plate frame, reseal them in the foil pouch, and return the foil pouch to 4°C.
3. Wash each well with 300 μ l of **1x** assay buffer. Allow to sit for at least 5 minutes. Discard the buffer, invert and blot dry plate. Do not let wells dry before proceeding to the next step.
4. Leave wells A-1 and A-2 empty as **Blank**.

Resistin (Mouse) ELISA Kit Protocol

5. Add 100µl of the prepared Mouse Resistin Standard solutions from #7 to #1 (reverse order of serial dilution) in duplicate to wells B-1 and B-2 to H-1 and H-2, respectively.
6. Add 100µl of Mouse Resistin positive control solution in duplicate.
7. Add 100µl diluted samples in duplicate into their designated wells.
8. Seal the immunoplate with acetate plate sealer (APS). Incubate for 2 hours at room temperature (20-23°C) on a plate shaker (300-400rpm).
9. Before washing the plate, remove the plate sealer carefully. Completely discard the liquid from wells. Wash each well with 300-350µl 1x assay buffer four times. At the end of the wash, discard the buffer, invert the plate and tap on a clean absorbent towel.
10. Add 100µl biotinylated anti-Mouse Resistin Detection Antibody into each well **except the Blank** well. Reseal the immunoplate with plate sealer and incubate for 2 hours at room temperature (20-23°C) on a plate shaker (300-400 rpm).
11. Wash 4 times with the 1x assay buffer as described in step 9.
12. Add 100µl SA-HRP solution into each well. Reseal the immunoplate with plate sealer and incubate the plate for 30 minutes at room temperature (20-23°C) on a plate shaker (300-400 rpm).
13. Wash 4 times with the 1x assay buffer as described in step 9.
14. Add 100µl substrate solution (TMB) provided in this kit into each well. Reseal the plate with plate sealer to protect from light and incubate the plate for 20-30 minutes at room temperature (20-23°C) on a plate shaker (300-400 rpm).
15. Add 100µl Stop Solution (2N Hydrochloric Acid) into each well to stop the reaction. The color in the well should change from blue to yellow. If the color change does not appear to be uniform, gently tap the plate to ensure thorough mixing. Go to the next step within 20 minutes.
16. Read the absorbance O.D. at 450 nm using a Microtiter Plate Reader.

Resistin (Mouse) ELISA Kit Protocol

ADDITIONAL RECOMMENDED PROCEDURAL NOTES:

- Reagents of different lot numbers should not be mixed.
- Recheck the reagent labels when loading the plate to ensure that everything is added correctly.
- Unused microplate strips should be placed back in the foil pouch with a dessicant and stored at 4°C. Do not allow moisture to enter the wells.
- When handling the plate, avoid touching the bottom.
- Manual washing may cause high duplicate coefficient variations. To reduce this factor, liquid from the plate should be removed by inverting and blotting the plate on an absorbent material.
- If the room temperature is not within the suggested range (20-23°C), variations in results may occur.
- The same reservoir for the reagents may be reused if the reservoir is washed well with distilled water before each use.
- Each laboratory must determine the appropriate dilution factors for the samples to be measured to ensure that the samples are within the dynamic range of the standard curve.
- High levels of interfering proteins may cause variations within the sample results; therefore, it is imperative to select the appropriate sample preparation procedure to obtain the optimal results.
- Each time a new tip is used, make sure the tip is secure and free of air bubbles. For better intra-assay variation, aspirate and expel a reagent or sample back into the container a few times prior to loading.
- Avoid submerging the whole tip into reagents because droplets can accumulate at the end of the tip causing an excess of reagent to be loaded into the well. This can lead to poor results.
- A multi-channel pipette is **NOT** recommended to load the biotinylated detection antibody or standard because variations in results may occur.
- For optimal results, an orbital plate shaker capable of 300-500 rpm is recommended for all incubations.
- Modification of the existing protocol (i.e. standard dilutions, pipetting technique, washing technique, incubation time or temperature, storage conditions, and kit expiration) may affect the sensitivity and specificity of the test.

Resitin (Mouse) ELISA Kit Protocol

SUMMARY OF ASSAY PROTOCOL

Add 100µl/well of Mouse Resistin Standard, Positive Control, or diluted Samples



Incubate at room temperature (20-23°C) for 2 hours



Wash immunoplate 4 times with 300-350µl/well of 1x assay buffer



Add 100µl/well Biotinylated anti-Mouse Resistin Detection Antibody



Incubate at room temperature (20-23°C) for 2 hours



Wash immunoplate 4 times with 300-350µl/well of 1x assay buffer



Add 100µl/well of SA-HRP solution



Incubate at room temperature (20-23°C) for 30 minutes



Wash immunoplate 4 times with 300-350µl/well of 1x assay buffer



Add 100µl/well of substrate solution (TMB)



Incubate at room temperature (20-23°C) for 20-30 minutes



Terminate the reaction with 100µl/well of Stop Solution



Read absorbance O.D. at 450nm and calculate results


Resistin (Mouse) ELISA Kit Protocol

CALCULATION OF RESULTS

Plot the standard curve on log-log graph paper. Known concentrations of Mouse Resistin Standard and its corresponding reading is plotted on the log scale (x-axis) and the log scale (y-axis) respectively. The standard curve shows a correlated relationship between Mouse Resistin concentrations and the corresponding O.D. absorbance. As the standard concentration increases, the intensity of the blue color increases, and in turn the O.D. absorbance, increases.

The concentration of Mouse Resistin in sample is determined by plotting the sample's O.D. on the Y-axis, then drawing a horizontal line to intersect with the standard curve. A vertical line dropped from this point will intersect the X-axis at a coordinate in the unknown sample.

Refer to QC Data sheet for acceptable values of the Positive Control.



Resistin (Mouse) ELISA Kit Protocol

STORAGE

1. Store the kit at 4°C upon receipt. The kit should be equilibrated to room temperature (20-23°C) before the assay.
2. Store **1x** assay buffer at 4°C.
3. Remove any unneeded strips from Mouse Resistin Antibody-Coated plate, reseal them in zip-lock foil pouch and keep at 4°C.
4. Keep rehydrated solution of Mouse Resistin Standard, Biotinylated Anti-Mouse Resistin Detection Antibody and SA-HRP at 4°C. Prepare only the required amount.

NOTE:

1. It is recommended that the solutions be used on the same day of rehydration.
2. Unextracted serum samples of normal subjects are to be diluted with **1x** assay buffer.
3. After adding stop buffer, read the plate within 20 minutes.

REFERENCES

1. Steppan, C.M. et al. 2001. The hormone resistin links obesity to diabetes. *Nature*. 409:307-312.
2. Steppan, C.M. et al. 2001. A family of tissue-specific resistin-like molecules. *Proc. Natl. Acad. Sci. U.S.A.* 98:502-506.
3. Way, J.M., et al. 2001. Adipose tissue resistin expression is severely suppressed in obesity and stimulated by peroxisome proliferator-activated receptor gamma agonists. *J. Biol. Chem.* 276:25651-25653.

Resistin (Mouse) ELISA Kit Protocol

ASSAY DIAGRAM

