

**Protocol for Catalog # FEK-069-04**  
**CCK (26-33) (Non-Sulfated)**  
**FLUORESCENT IMMUNOASSAY KIT**  
**(range: 0-10,000pg/ml)**




PHOENIX PHARMACEUTICALS, INC.

# INTRODUCTION

*This Fluorescent Immunoassay kit is designed to detect a specific peptide and its related peptides based on the principle of “competitive” enzyme immunoassay.*

## PRINCIPLE OF ENZYME IMMUNOASSAY WITH THIS KIT

The immunoplate in this kit is pre-coated with a secondary antibody and the nonspecific binding sites are blocked. The secondary antibody can bind to the Fc fragment of the primary antibody (peptide antibody) whose Fab fragment will be competitively bound by both biotinylated peptide and peptide standard or targeted peptide in samples. The biotinylated peptide interacts with streptavidin-horseradish peroxidase (SA-HRP) which catalyzes the substrate. The fluorescence intensity is directly proportional to the amount of biotinylated peptide-SA-HRP complex but inversely proportional to the amount of the peptide in standard solutions or samples. This is due to the competitive binding of the biotinylated peptide with the standard peptide or samples to the peptide antibody (primary antibody). A standard curve of known concentration can be established accordingly. The unknown concentration in samples can be determined by extrapolation to this standard curve.



CAUTION: INVESTIGATIONAL DEVICE. LIMITED BY LAW TO INVESTIGATIONAL USE.  
FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

## KIT MATERIALS

1. 20x assay buffer concentrate (50ml)
2. 96 well immunoplate (1 plate)
3. Acetate plate sealer (APS), (3 pieces)
4. Primary antibody (rabbit anti-peptide IgG) (1 vial)
5. Standard peptide (1 µg)
6. Biotinylated peptide (1 vial)
7. Streptavidin-horseradish peroxidase (SA-HRP) (30µl)
8. Substrate solution (13ml)
9. Positive control (1 vial)
10. Stable peroxide solution (1.5ml)
11. Stop solution (13ml)
12. Assay diagram (1 sheet)
13. General protocol (1 book)

NOTE: *Phoenix Pharmaceuticals, Inc. guarantees that its products conform to the information contained in this publication. The purchaser must determine the suitability of the product for their particular needs and establish optimum sample concentration.*

## MATERIALS REQUIRED BUT NOT SUPPLIED

1. Fluorescence microplate reader (325nm to 420nm)
2. Orbital plate shaker capable of 300-400rpm (recommended)
3. Microtiter plate washer (recommended)
4. Multi-channel pipette capable of dispensing 50-100µl (recommended)
5. Solution reservoir (recommended)
6. Absorbent material for blotting.
7. Lavender Vacutainer blood collection tubes (optional)
8. Aprotinin (0.6TIU/ml of blood) (optional) ..... **Catalog no. RK-APRO**
9. Buffer A (optional) ..... **Catalog no. RK-BA-1**
10. Buffer B (optional)..... **Catalog no. RK-BB-1**
11. C18 SEP-COLUMN (optional)..... **Catalog no. RK-SEPCOL-1**


**NOTE:** *The kit should be equilibrated to room temperature (20-23°C) before opening any vials and starting the assay. It is highly recommended that the solutions be used as soon as possible after rehydration. Recommended blood collection protocol is provided on page 9. Each kit contains sufficient reagents for 96 wells and is capable of assaying 40 duplicate samples.*

## **ASSAY PROCEDURE**

1. Thoroughly read this protocol before performing an assay. Please allow all kit components to return to room temperature before opening packages (25-45 minutes).
2. Dilute the **20x** assay buffer concentrate with 950ml of distilled water. This will be the **1x** assay buffer solution used to dilute or reconstitute all other reagents in this kit and samples.  
**Note:** If crystals appear in the **20x** assay buffer, the buffer can be placed in a warm water bath for approximately 30 minutes or until no crystals are visible. Mix thoroughly before use.
3. Centrifuge and dilute the standard peptide with 1ml of **1x** assay buffer, vortex. The concentration of this stock solution is 1,000ng/ml. Allow the solution to sit at least 10 minutes at room temperature (20-23°C) to completely dissolve in solution. Centrifuge and vortex immediately before use.

### **Prepare peptide standard solutions as follows:**

<b>Standard No.</b>	<b>Std. volume</b>	<b>1x Assay Buffer</b>	<b>Concentrations</b>
Stock	1000µl	----	1,000 ng/ml
#1	100µl of stock	900µl	100 ng/ml
#2	100µl of #1	900µl	10,000 pg/ml
#3	100µl of #2	900µl	1,000 pg/ml
#4	100µl of #3	900µl	100 pg/ml
#5	100µl of #4	900µl	10 pg/ml
#6	100µl of #5	900µl	1 pg/ml



4. Rehydrate primary antibody with **1x** assay buffer. Refer to **QC data sheet for instructions on rehydrating the primary antibody**. Allow to sit for at least 5 minutes to completely dissolve. Mix thoroughly.
5. Rehydrate the positive control with **1x** assay buffer. Refer to **QC data sheet for instructions on rehydrating the positive control**.
6. Leave wells A-1 and A-2 empty as **Blank**.
7. Add 50µl of **1x** assay buffer into wells B-1 and B-2 as **Total Binding**.
8. Add 50µl of prepared peptide standards from **#6 to #2** (in reverse order of serial dilution) into wells from C-1 and C-2 to G-1 and G-2 respectively.  
*Note: Peptide standards should be assayed in duplicate.*
9. Add 50µl of positive control solutions in wells H1-H2.
10. Add 50µl of prepared samples in their designated wells in duplicate.
11. Add 25µl of rehydrated primary antibody into each well **except** the **Blank** well. Gently tap the plate to ensure thorough mixing.  
*Note: A multi-channel pipette is **NOT** recommended to load the primary antibody because variations in results may occur.*
12. Seal the immunoplate with the acetate plate sealer (APS). Incubate the immunoplate overnight (approximately 16-24 hours) at 4°C.
13. The next day, rehydrate the biotinylated peptide with **1x** assay buffer. Refer to **QC data sheet for instructions on rehydrating the biotinylated peptide**. Allow to sit for at least 5 minutes to completely dissolve. Mix thoroughly. Remove APS from immunoplate. **DO NOT WASH THE IMMUNOPLATE**. Add 25µl of rehydrated biotinylated peptide into each well except the **Blank** well.

*Note: A multi-channel pipette is **NOT** recommended to load the biotinylated peptide because variations in results may occur.*

14. Seal the immunoplate with acetate plate sealer (APS). Incubate the immunoplate for 1.5 hours at room temperature (20-23°C). Orbital shaking at 300-400 rpm is recommended for the duration of the incubation.
15. Centrifuge the SA-HRP vial provided in this kit (3,000-5,000 rpm, 5 seconds) and pipette 12µl of SA-HRP into 12ml of **1x** assay buffer to make SA-HRP solution, vortex thoroughly.
16. Remove APS from the immunoplate. Discard contents of the wells.
17. Wash each well with 350µl of **1x** assay buffer, discard the buffer, invert and blot dry plate. Repeat 4 times.
18. Add 100µl of SA-HRP solution into each well.
19. Reseal the immunoplate with APS. Incubate for 1 hour at room temperature (20-23°C). Orbital shaking at 300-400 rpm is recommended for the duration of the incubation.
20. Mix 9 parts of Substrate Solution to 1 part of Stable Peroxide Solution. This working solution is stable for 24 hours at room temperature (20-23°C) and no protection from light is required. To reduce variability, equilibrate the working solution to room temperature (20-23°C) before adding to the wells.
21. Remove APS from the immunoplate.
22. Wash and blot dry the immunoplate 4 times with **1x** assay buffer as described above in step 17.
23. Add 100µl of prepared Substrate Solution into each well. Gently tap the immunoplate to ensure thorough mixing.
24. Reseal the immunoplate with APS. Incubate for 15-20 minutes at room temperature (20-23°C). Orbital shaking at 300-400 rpm is recommended for the duration of the incubation.
25. Remove APS from the immunoplate. Add 100µl of stop solution into each well to stop the reaction. Gently tap the plate to ensure thorough mixing. Proceed to the next step within 20 minutes.
26. Load the immunoplate onto a fluorescence microplate reader, and measure relative fluorescence units (RFU) of each well. The excitation and emission maxima for the Substrate Solution are 325 and 420 nm, respectively. Wavelengths between 315 and 340 nm for excitation and 370 and 470 nm for emission also can be used for detection.

## Additional Recommended Procedural Notes:

- Each laboratory must determine the appropriate dilution factors for the samples to be measured to ensure that the samples are within the dynamic range of the standard curve.
- High levels of interfering proteins may cause variations within sample results; therefore, it is imperative to select the appropriate sample preparation procedure to obtain optimal results.
- Reagents of different lot numbers should not be mixed.
- Recheck the reagent labels when loading the plate to ensure that everything is added correctly.
- Manual washing may cause high duplicate coefficient variations. To reduce this factor, liquid from the plate should be removed by inverting and blotting the plate on an absorbent material.
- If the room temperature is not within the suggested range (20-23°C), variations in results may occur.
- The same reservoir for the reagents may be reused if the reservoir is washed well with distilled water before each use.
- Each time a new tip is used, make sure the tip is secure and free of air bubbles. For better intra-assay variation, aspirate and expel a reagent or sample back into the container a few times prior to loading.
- Avoid submerging the whole tip into reagents because droplets can accumulate at the end of the tip causing an excess of reagent to be loaded into the well. This can lead to poor results.
- For optimal results, an orbital plate shaker capable of 300-400 rpm is recommended for all incubations.
- Modification of the existing protocol (i.e. standard dilutions, pipetting technique, washing technique, incubation time or temperature, storage conditions, and kit expiration) may affect the sensitivity and specificity of the test.
- Fluorometric units are typically defined as relative fluorescence units (RFU) because the integrated signal is dependent on instrument settings. Consult the fluorometer's user manual for specific instrument capabilities and settings.

## Calculation of results

Plot the standard curve on semi-log graph paper. A standard curve is constructed by plotting the known concentrations of standard peptide on the log scale (X-axis), and its corresponding fluorescence intensity reading on the linear scale (Y-axis). It is strongly recommended to use curve-fitting software capable of 4 parameter logistics or log-logit to quantify the concentration of standard peptide. The standard curve shows an inverse relationship between peptide concentrations and the corresponding fluorescence intensity. As the standard concentration increases, the fluorescence intensity decreases.

The concentration of peptide in a sample is determined by locating the sample's O.D. on the Y-axis, then drawing a horizontal line to intersect with the standard curve. A vertical line drawn from this point will intersect the X-axis at a coordinate corresponding to the peptide concentration in the sample. If samples have been diluted prior to the assay, the measured concentration must be multiplied by their respective dilution factors.

The standard curve will be a reverse sigmoidal shape.

Refer to QC Data Sheet for acceptable values of the positive controls.

## **STORAGE**

1. Store the kit at 4°C upon receipt.
2. It is highly recommended that solutions be used as soon as possible after rehydration.
3. Store 1x assay buffer at 4°C.
4. Keep rehydrated solution of Standard, Biotinylated peptide, Antibody and HRP at 4°C.



## SUMMARY OF ASSAY PROTOCOL

Add 50 $\mu$ l/well of standard, positive control, or sample, and 25 $\mu$ l primary antibody

Incubate overnight (approximately 16-24 hours) at 4°C

Add 25 $\mu$ l/well of biotinylated peptide

Incubate at room temperature (20-23°C) for 1.5 hours

Wash immunoplate 4 times with 350 $\mu$ l/well of **1x** assay buffer

Add 100 $\mu$ l/well of SA-HRP solution

Incubate at room temperature (20-23°C) for 1 hour

Wash immunoplate 4 times with 350 $\mu$ l/well of **1x** assay buffer

Add 100 $\mu$ l/well of prepared substrate solution

Incubate at room temperature (20-23°C) for 15-20 minutes

Terminate reaction with 100 $\mu$ l/well of stop solution

Measure relative fluorescence intensity of each well and calculate results

## SUGGESTED METHOD FOR THE EXTRACTION OF PEPTIDES FROM PLASMA

---

### Blood Withdrawal:

Collect blood samples into the Lavender Vacutainer tubes which contain anti-coagulant and can collect up to 7ml of blood. Gently rock the Lavender Vacutainer tubes several times immediately after collection of blood for anti-coagulation. Transfer the blood from the Lavender Vacutainer tubes to the centrifuge tubes containing aprotinin (0.6 TIU /ml of blood) and gently rock several times to inhibit the activity of proteinases. Centrifuge the blood at 1,600 x g for 15 minutes at 4°C and collect the plasma. Plasma kept at -70°C may be stable for one month. If Lavender Vacutainer tubes are centrifuge-safe, the aprotinin may be added into the initial collection step.

### Extraction of Peptides from Plasma:

1. Acidify the plasma with an equal amount of buffer A. For example, if you are using 1ml of plasma, add 1ml of buffer A. Mix and centrifuge at 6,000 to 17,000 x g for 20 minutes at 4°C.
2. Equilibrate a SEP-COLUMN containing 200mg of C18 (Cat. No. RK-SEPCOL-1) by washing with buffer B (1ml, once) followed by buffer A (3ml, 3 times).

**Note:** From steps 3-5, no pressure should be applied to the column.

3. Load the acidified plasma solution onto the pre-equilibrated C-18 SEP-Column.
4. Slowly wash the column with buffer A (3ml, twice) and discard the wash.
5. Elute the peptide slowly with buffer B (3ml, once) and collect the eluant into a polystyrene tube.
6. Evaporate eluant to dryness in a centrifugal concentrator or by a suitable substitute method.

7. Keep the dried extract at -20°C and perform the assay as soon as possible. Use assay buffer to reconstitute the dried extract. If the peptide value does not fall within the range of detection, dilute or concentrate the sample accordingly.

### **Tips for extraction of plasma:**

When using a C-18 SEP-COLUMN for the first time, use a bulb (not supplied) to apply pressure to the column after the addition of 1ml of buffer B to facilitate the flow through the column. From steps 3-5, no pressure should be applied.

Ensure there is a constant flow for all solutions during the extraction procedure. Do not allow air bubbles to enter the C-18 matrix for optimal sample processing and recovery.

### **Drying Sample After Extraction:**

A combination of a centrifugal concentrator (i.e. Speedvac) and a lyophilizer (freeze-dryer) produces the best results for drying the sample after extraction. First, use a Speedvac to dry sample for approximately 15 minutes to remove the organic layer. Then snap-freeze the remaining sample, and freeze-dry overnight using a lyophilizer. This two-step procedure produces a more consistent fluffy powder that is easier to rehydrate than a sample dried only with a centrifugal concentrator. However, if a centrifugal concentrator is not accessible, freeze-drying overnight using a lyophilizer will be sufficient.

## **REFERENCES**

1. Porstmann, T. and Kiessig, S.T., Enzyme Immunoassay Techniques, An Overview, *Journal of Immunological Methods*, 150: 5-21 (1992).
2. Avrameas, S., Amplification Systems in Immunoenzymatic Techniques, *Journal of Immunological Methods*, 150: 23-32 (1992).

# ASSAY DIAGRAM



## USA

### PHOENIX PHARMACEUTICALS, INC.

330 Beach Rd. Burlingame,  
California 94010  
Tel: 650-558-8898, 800-988-1205  
Fax: 650-558-1686  
Info@PhoenixPeptide.com  
www.PhoenixPeptide.com

## Europe

### PHOENIX EUROPE GMBH

Viktoriastrasse 3-5  
D-76133 Karlsruhe  
Germany  
Tel: +49(721)1611950  
Fax: +49(721)1611952  
Germany@PhoenixPeptide.com